

WBNA – License to Cover for CP (LMS#0000035747)

SPECIAL CONDITIONS:

The above captioned Construction Permit contained the following special condition:

“The grant of this construction permit is subject to the condition that, with ample time before commencing operation, you make a good faith effort to identify and notify health care facilities (e.g., hospitals, nursing homes, see 47 CFR 15.242(a)(1)) within your service area potentially affected by your DTV operations. Contact with state and/or local hospital associations and local governmental health care licensing authorities may prove helpful in this process. During this pre-broadcast period, you must provide all notified entities with relevant technical details of your operation, such as DTV channel, targeted on-air date, effective radiated power, antenna location, and antenna height. You are required to place in the stations public inspection file documentation of the notifications and contacts made and you may not commence operations until good faith efforts have been made to notify affected health care facilities. During this pre-broadcast period and for up to twenty (20) days after commencing operations, should you become aware of any instances of medical devices malfunctioning or that such devices are likely to malfunction due to your DTV operations, you must cooperate with the health care facility so that it is afforded a reasonable opportunity to resolve the interference problem. At such time as all provisions of this condition have been fulfilled, and either upon the expiration of twenty (20) days following commencement of operations or when all known interference problems have been resolved, whichever is later, this condition lapses.”

Licensee has prepared and mailed to eligible entities a notification as required by the above condition and has placed a copy of the form of letter sent as well as a mailing list of the parties to which it was sent in the WBNA Online Public File in the Applications & Related Materials – Additional Documents – License Related Documents subfolder.

Licensee will monitor for communications from such notified parties of any claims of medical device interference and will cooperate with any affected party to resolve the interference issue.